




Abteilung für Medizinische Strahlenphysik
Inselspital, Universität Bern, Schweiz
www.ams.unibe.ch



Production and Properties of X Rays (2)

F. Hasenbalg
Abteilung für Medizinische Strahlenphysik
Inselspital - Universität Bern

e-mail: hasenbalg@ams.unibe.ch



Results assignment 05: X-Ray Production



Gnuplot:
plot '100kv.spc' smooth csplines w lin 1

Spectrum mean energies (weighted averages)

Anode

100 kV 28.46 keV


125 kV 33.59 keV

After 3 mm Al

100 kV 48.34 keV

125 kV 53.55 keV

Quality of X Rays: Beam Hardening



Beam Hardening

Larger absorber thickness
 -> fewer low energy photons
 -> higher mean beam energy

Determination through **homogeneity grade H**

$$H = \frac{d_{1/2}^{(1)}}{d_{1/2}^{(2)}} \quad d_{1/2}: \text{half value layer}$$

homogeneous beam **H = 1**

heterogeneous beam **H < 1**

Dif. energy fluence: $E \frac{d\Phi}{dE}$ shifts the distribution towards the high energy end of the spectrum and the area under the curve gives the total energy fluence.

Disadvantages of low energy photons:

Diagnostic: they do not reach the imaging device but contribute to the dose to the patient

Therapy (superficial tumors): do not penetrate to the tumor and increase the dose to the skin


Half value layer: measures the beam ability to penetrate some known material

Al: U < 120 kV

Cu: 120- 400 kV

Quality of X Rays: Beam Hardening

50 keV Photons in Aluminum



Previous example with a **homogeneous** 50 keV photon beam in aluminum


$$H = 1$$

Histogram analysis





New example with a **heterogeneous** 100 kV photon beam in aluminum

$$H < 1$$

100 kV X Rays



Beam Hardening



EGS-Ray: attenuation of a heterogeneous photon beam

pencil beam (parallel geometry)

Program control:


Generation

Only the doses of a determined generation (primary, secondary, etc) of particles is registered (number of the generation to be summed)
Generation
1
adds the doses of only the first interacting particles, primary photons in this case. For more details look in the manual.

Histogram analysis on x or y direction.

Result and performance independent of voxel size.

```
Daten
D:\...\521icru.dat
Randomseed
1802 9373
Histories
100000
Punktquelle
0 0 0
Richtung
0 0 1
Photonen
Spektrum
D:\...\100kv.spc
Rechenraum
-4 -4 0 4 4 15
Scoringraum
-1 -1 10 1 1 15
Voxelgröße
2 2 1
PLatte
AL521ICRU
2 10 15
Histogramm
2 10
Histogramm
2 10.01
Histogramm
2 10.02
...
Histogramm
2 14
Generation
1
```



Suggested values: 0.1, 0.2, 0.5, 07,
1, 2, ... 10,
12.5, 15, 17.5, 40 mm

EGS-Ray: attenuation of a heterogeneous photon beam (2)

Anode tension: 100 kV

Al filter

Pb collimator

divergent beam

Histogram analysis

Daten

D:\... \521icru.dat

Randomseed

1802 9373

Histories

400000

Punktquelle

0 0 0

Richtungsbereich

-1 1 -1 1 0.99 1

Photonen

Spektrum

D:\... \100kV.spc

Rechenraum

-2 -2 -0.5 2 2 7

Scoringraum

-0.5 -0.5 2 0.5 0.5 7

Voxelgröße

1 1 1

photon
source

Anode
spectrum
100 kV

Halbraum

AIR521ICRU

2 2 1

Histogramm

2 2

rem Filter

rem 10 mm

PLatte


AL521ICRU 10 mm Al filter

2 2 3

Histogramm

2 6.05

Presta



Histogram to
check backscatter

Al filter

Anode photon source


Vacuum

Air

AIR521ICRU
ECUT = 0.0 MeV
PCUT = 0.0 MeV
Dichte = 0 g/ccm

1 cm

Ignore
Protokoll
messages !



Results are voxel independent

EGS-Ray: attenuation of a heterogeneous photon beam (3)


Anode tension: 100 kV

3mm Al filter

divergent beam

**Total doses
addition**

	Daten	
	D:\...\521icru.dat	
	Randomseed	
	1802 9373	
	Histories	
	100000	
	Punktquelle	
	0 0 0	
photon	Richtungsbereich	
source	-1 1 -1 1 0.999 1	
	Photonen	
Anode	Spektrum	
spectrum	D:\...\100kV.spc	
100 kV	Rechenraum	
	-5 -5 -2 5 5 114	
	Scoringraum	
	-5 -5 -1 5 5 114	
	Voxelgröße	
	2 2 2	
	PLatte	
	AL521ICRU	3 mm
	2 3 3.3	Al filter
	Histogramm	
	2 3.5	
	PLatte	
	AL521ICRU	10 mm
	2 10 11	additional
	Histogramm	Al filter
	2 50	
	Box	
	PB521ICRU	
1 cm ³	-0.5 -0.5 100 0.5 0.5 101	
Pb detector	Presta	



Use
-> Objekte
-> integrale
Energie dosis
of Pb box
to make the
measurements

Assignment 06: X ray properties

X-ray properties:

Use one of the anode spectra obtained in part 1) either 100 kV or 125 kV (or both)

Simulate: **parallel** X ray beams

with several filters: 0.1, 0.2, 1, 2, ..., 10, 12.5, 15, 17.5 and 30 mm Al

Analysis: Using **histogram analysis** determine the 1st and 2nd half value layers and the homogeneity grade:

$$H = \text{HVL1}/\text{HVL2}$$

Discuss the limitations of this approach for real measurements

Simulate: a) **divergent** X ray beams

with several filters: 0.1, 0.2, 1, 2, ..., 10, 12.5, 15, 17.5 and 30 mm Al

Measure the photon intensity using a histogram analysis

b) **divergent** X ray beams + 3mm Al filter

with several additional filters: 1, 3, 7, 5, 11, 13 and 15 mm Al

Measure the photon intensity with a “virtual” ionization chamber

Source to detector distance: 100 cm

Analysis: Using **histogram analysis** determine the 1st and 2nd half value layers and the homogeneity grade:

$$H = \text{HVL1}/\text{HVL2}$$

Discuss the results

Scoring dimensions and voxel size adapted to each case. Enough statistics.